

स्वामी रामानंद तीर्थ ग्रामीण शासकीय वैद्यकीय महाविद्यालय अंबाजोगाई

(अधिष्ठाता कार्यालय)

SWAMI RAMANAND TEERTH RURAL GOVERNMENT MEDICAL COLLEGE, AMBAJOGAI DIST BEED
PH NO (02446) 245792/2449362/248438/- D. FAX- 02446/247132 Website www.srtrmca.org email-srt.acad@rediffmail.com
SRTRGMCA/STORE/20-21/PLAN/337/2021 DATE 15/01/2021

Sub: - Quotation for Purchase of Instrument & Equipment.....

We are inviting the quotations as per G.R. for the following machinery and equipment's required at different department of this hospital. Quotation should be in the prescribed form and as per Terms and condition as follows.

Please quote your competitive rates in Indian Rupees only for the following items. Rates should be within the market rate Limits.

TERMS & CONDITIONS


1. GST Registration Certificate.
2. Authorization Letter: - A dealer quoting on behalf of manufacturer must attach authorization letter from the manufacturer to self the equipment & to do after sales services as per.
3. ISO OR ISI Certificates: - In case of equipment with ISO or ISI mark, as per Quotation requirement relevant certificates from ISO or ISI should be attached.
4. Product Catalogue: - Original Catalogues / brochures should be enclosed giving technical details of the offered equipment. All specification should be substantiated by data sheets in original.
5. The delivery of goods should invariably by make with 6 weeks for indigenous terms and 12 weeks for imported items, from the date of confirm order. If the date of delivery cannot be strictly adhered to, then extension required by the quotation should be stated in the bid. No excuses for delay by any statutory like custom etc. Will be taken into consideration for extension of the period of delivery. The primary responsibility for supply of goods in time will be rest with the supplier.
6. Penalty on late delivery :- In the event of the late delivery of goods, the purchaser will recover for contractor by way penalty sum equal to half percent of price of the good (subject to maximum 5%) delivered late per week calculated from the next day after the agreed delivery period is over.
7. Warranty Period :- (including supply of spares) the warranty period shall be for Two years from the date of commissioning of all equipment's supplied as certified by the consignee.
8. Replacement of parts and service thereof due to manufacturing defects during warranty period will be entirely at the suppliers cost. The expenditure incurred for procurement of space transport, installation, commissioning and various duties there of shall be borne by the supplier.
9. Annual Maintenance Contract/ Comprehensive Maintenance Contract: - The supplier will have to agree to enter into a annual maintenance contract (AMC) @ 1% of the order value of the equipment Where required quotation will have to agree for comprehensive maintenance (CMC)

inclusive of all spares @ 5% of the order value of the equipment. The period of such AMC/CMC will be of 8 years after completion of warranty period, in case of non-communication of AMC/CMC the supplier will be liable to pay a Penalty. Such penalty shall be recovered from the amount of bank guarantee. Payment for AMC/CMC on yearly basis will be made by the user institution, at the end of year after satisfactory performance report from the end user.

10. Payment: - Install payment of 90% of the contract value will be made on delivery and installation. Remaining 10% of contract value will be paid only on satisfactory commissioning of the machinery within 30 days or 1 month.
11. The successful quotation must ensure 100% uptime during warranty period in case of down time; warranty period will be extended for period of down time. The quotation will have to give Bank Guarantee (of Nationalized Bank) of 5% of ordered value valid for 10 years towards performance guarantee for warranty. If the equipment is not attended within 24 hours for Ambajogai and 48 hours for other places the supplier will be available to pay a penalty of 0.05% of purchase cost forever day of delay. Such penalty shall be recovered from the amount of bank guarantee; Certificate of such uptime/downtime shall be issued by the end user. Which shall be binding for the supplier?
12. Machinery Specification enclosed here with. The instrument supplied should be as per specifications.

Note:-THE DECISION OF THE DEAN, SWAMI RAMANAND TEERTH RURAL GOVERNMENT MEDICAL COLLEGE & HOSPITAL AMBAJOGAI DIST BEED RESERVES THE RIGHT TO REJECT WITHOUT ASSIGNING ANY REASON.

Please superscript the envelope with "QUOTATION FOR PURCHASE OF -----"due date.
Sealed quotation should reach this office on or before date: - 22 /01/2021.


Dean,
Swami Ramanand Teerth Rural Govt,
Medical College, Ambajogai, Dist-Beed.

To,


- 1) -----
- 2) -----
- 3) -----
- 4) Notice Board & Website Development Committee, S.R.T.R.G. Medical College, Ambajogai.

Specification of Instrument & equipment (Govt.)

| Sr. No. | Item Name | Approx. Qty Required | Specification |
|--|------------------------------|----------------------|---|
| Dept. of ENT (कान, नाक व घसा शास्त्र) | | | |
| 1 | 30 degree Nasal Endoscope | 01 | <ol style="list-style-type: none"> 1.Outer diameter 4 mm diameter. 2.Length should be between 180-185 mm 3.Vision 30 degree with 105 degree wide angle 4.Fibreoptic light transmission incorporated 5.It should offer full O ring illumination 6.High depth of focu high speed round lens with high magnification with sapphire glass cover. 7.Stainless steel body 8.Tube laser welded to body welded 9.The scope should be autoclavable 10.Low vignetty 11.The scope should be autoclavable 12.Should have USFDA certification. |
| 2 | 45 degree Nasal Endoscope | 01 | <ol style="list-style-type: none"> 1.Outer diameter 4 mm diameter 2.Lenght should be between 180-185 mm 3.Vision 45 degree with 105 degree wide angle 4.Fibreoptic light transmission incorporated 5.It should offer full O ring illumination. 6.High depth of focu high speed round lens with high mangnification with sapphire glass cover. 7. Stainless steel body. 8.Tube laser welded to body 9.The scope should be autoclavable 10.Low vignetty 11.The scope should be autoclavable 12.Should have USFDA certification. |
| 3 | Single chip Endoscopy system | 01 | <ul style="list-style-type: none"> •Image sensor:1/2' CCD-Chip&more •Resolution:>450 lines (Horizontal) •Pixels:752 (H)x582 (V) •Signal-to -noise ration:>60 dB •AGC: Microprocessor controlled •Lense: C Mount Connection Weight:75g •Min. sensitivity:3 Lux (f 1.4) •Dimensions: Diameter 23-33 mm, length 53mm. •Certified to IEC 601-2-18, CSA 22.2 No 601] UL 2601 and CE acc to •MDD Protection class1/BF •2 programmable function keys on the camera head for control of |

| | | | |
|--|----------------------------------|----|--|
| | | | <ul style="list-style-type: none"> •camera functions or video printer/recorder functions & other peripheral units. •Adaptable to an operating microscope by means of a special quick adapter •Camera head fully soakable for disinfection |
| 4 | Single Chip Endoscopy Camera CCU | 01 | <ul style="list-style-type: none"> •Should be fully compatible to the color systems PAL&NTSC. All the necessary setting required to be done during surgery in the camera should be under sterile area -Digital contrast enhancement. -Digital anti-moire/anti-grid filter for use with fiberscopes. •Should have the Auto Exposure System for consistent illumination level. •High Horizontal image resolution of more than 840 lines, therefore even the finest variations in tissue structures are perceivable on a high- resolution monitor. •Automatic white balance with memory function for two settings. •Multilingual menu system. •Vdieo Output: Composite signal to BNC socket, S-Video signal to pin Mini-DIN socket 2x •World power supply Certified to IEC 609-1, 609-2-14, CSA 22.2 No 609, UL 2609 and CE acc to MDD, protection class I/BF |
| Dept. of FMT (न्यायवैद्यकशास्त्र) | | | |
| 5 | Fogger Machine | 03 | <ol style="list-style-type: none"> 1.Device should be made for hospital use only. (operation theatre ot intensive care unit sterilization) 2.Device should be portable, electrically operated. 3.Machine weight at least 6 kg 4.Tank capacity at around 7 liters. 5Area covered should be around 1000 sq feet. 6.Tank should be metallic made and sturdy. 7.Droplet size generated by nozzle should be submicron size. 8.Timer function should be present. Integrated timer is preferred. 9.Particle throw in a closed room should reach up to 3 to 4 meters. 10.Output should be compatible with acidic or alkaline liquids. 11.All parts should be compatible with acidic or alkaline liquids. 12.Should have intake air filter uniquely designed of two layes for dust and fog separation 13.Device motor should be high speed. 14.Should have precision metering System:0-70ml/minute 15.Should be of nozzle assembly- non clogging desing- Engg. Plastics. <p><u>Specification</u></p> <ul style="list-style-type: none"> •Power supply AC 220 22V, 051 Hz •Max negative pressure:>0.09Mpa •Noise:<60 DB (A) •Input power:<180VA •Flow Rate:>20 L/min •Jar capcity: 2500ml/PC, 2 pieces |

| | | | |
|---|-----------|----|---|
| 6 | Bone Sets | 03 | <ul style="list-style-type: none"> •Light weight •Crack proof •Excellent finish •20& bones, one set of male and another set of female skeleton. •must closely resemble original bones in evey way. |
|---|-----------|----|---|


 Dean
 Swami Ramanand Teerth Govt.
 Medical College, Ambajogai Dist. Beed