



स्वामी रामानंद तीर्थ ग्रामीण शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, अंबाजोगाई.  
(अधिष्ठाता यांचे कार्यालय)

SWAMI RAMANAND TEERTH RURAL GOVT. MEDICAL COLLEGE & HOSPITAL, AMBAJOGAI DIST BEED  
PH NO (02446)245792/248963/249362/248438/-D, FAX- 02446/247132 Website www.srtrmca.org email-srt.acad@rediffmail.com

NO.SRTRGMC&H'A/STORE/ PLAN/ 194 /2018

DATE- / /2018

9/1/19

**Sub: - Quotation for Purchase of Instrument & Equipment.....**

We are inviting the quotations as per G.R. for the following machinery and equipment's required at different department of this hospital. Quotation should be in the prescribed form and as per Terms and condition as follows.

Please quote your competitive rates in Indian Rupees only for the following items. Rates should be within the market rate Limits.

**TERMS & CONDITIONS**

1. Sales tax / Vat Clearance Certificate: - Attested copy of Sales Tax VAT Clearance Certificate mentioning clearance up to Oct 2016 or wherever applicable sales tax exemption certificate form sales tax authority.
2. Authorization Letter: - A dealer quoting on behalf of manufacturer must attach authorization letter from the manufacturer to self the equipment & to do after sales services as per.
3. ISO OR ISI Certificates: - In case of equipment with ISO or ISI mark, as per Quotation requirement relevant certificates from ISO or ISI should be attached.
4. Product Catalogue: - Original Catalogues / brochures should be enclosed giving technical details of the offered equipment. All specification should be substantiated by data sheets in original.
5. The delivery of goods should invariably be made with 6 weeks for indigenous terms and 12 weeks for imported items, from the date of confirm order. If the date of delivery cannot be strictly adhered to, then extension required by the quotation should be stated in the bid. No excuses for delay by any statutory like custom etc. Will be taken into consideration for extension of the period of delivery. The primary responsibility for supply of goods in time will be rest with the supplier.
6. Penalty on late delivery :- In the event of the late delivery of goods, the purchaser will recover for contractor by way penalty sum equal to half percent of price of the good (subject to maximum 5%) delivered late per week calculated from the next day after the agreed delivery period is over.
7. Warranty Period :- (including supply of spares) the warranty period shall be for Two years from the date of commissioning of all equipment's supplied as certified by the consignee.
8. Replacement of parts and service thereof due to manufacturing defects during warranty period will be entirely at the suppliers cost. The expenditure incurred


for procurement of space transport, installation, commissioning and various duties there of shall be borne by the supplier.

9. Annual Maintenance Contract/ Comprehensive Maintenance Contract: - The supplier will have to agree to enter into a annual maintenance contract (AMC) @ 1% of the order value of the equipment Where required quotation will have to agree for comprehensive maintenance (CMC) inclusive of all spares @ 5% of the order value of the equipment. The period of such AMC/CMC will be of 8 years after completion of warranty period, in case of non-communication of AMC/CMC the supplier will be liable to pay a Penalty. Such penalty shall be recovered from the amount of bank guarantee. Payment for AMC/CMC on yearly basis will be made by the user institution, at the end of year after satisfactory performance report from the end user.
10. Payment: - Install payment of 90% of the contract value will be made on delivery and installation. Remaining 10% of contract value will be paid only on satisfactory commissioning of the machinery within 30 days or 1 month.
11. The successful quotation must ensure 100% uptime during warranty period in case of down time; warranty period will be extended for period of down time. The quotation will have to give Bank Guarantee (of Nationalized Bank) of 5% of ordered value valid for 10 years towards performance guarantee for warranty. If the equipment is not attended within 24 hours for Ambajogai and 48 hours for other places the supplier will be available to pay a penalty of 0.05% of purchase cost forever day of delay. Such penalty shall be recovered from the amount of bank guarantee; Certificate of such uptime/downtime shall be issued by the end user. Which shall be binding for the supplier?
12. Machinery Specification enclosed here with.

Note:-THE DECISION OF THE DEAN, SWAMI RAMANAND TEERTH RURAL GOVERNMENT MEDICAL COLLEGE & HOSPITAL AMBAJOGAI DIST BEED RESERVES THE RIGHT TO REJECT A WITHOUT ASSIGNING ANY REASON.

Please superscript the envelop with "QUOTATION FOR PURCHASE OF -----  
-----"due date.

Sealed quotation should reach this office on or before date: - 19 /01/2018.

  
Dean,

Swami Ramanand Teerth Rural Govt,  
Medical College & Hospital, Ambajogai, Dist-Beed.

To,

- 1) -----
- 2) -----
- 3) -----
- 4) Notice Board & Website Development Committee, S.R.T.R.G. Medical College & Hospital, Ambajogai.

## Specification of Instrument & equipment (PLAN)

Sr. No.	Item Name	Approx Qty Required	Specification
<b>Dept Of ENT (कान नाक व घसा शास्त्र )</b>			
1	Endoscope 0 Degree	01	<p>Technical Specification</p> <p>Specification for O degree endoscope ( sinuscope)</p> <ol style="list-style-type: none"> <li>1. Outer diameter 4 mm diameter</li> <li>2. Length should be between 180-185 mm</li> <li>3. Vision 30 degree with 105 degree wide angle</li> <li>4. Fibreoptic light transmission incorporated</li> <li>5. It should offer full O ring illumination</li> <li>6. High depth of focu high speed round lens with high magnification with sapphire glass cover</li> <li>7. Stianless steel body</li> <li>8. Tube laser welded to body welded</li> <li>9. The scope should be autoclavable</li> <li>10. The scope should be autoclavable</li> <li>11. Low vignetty</li> </ol>
2	Endoscope 3 0 Degree	01	<p>Technical Specification</p> <p>Specification for 3O degree endoscope ( sinuscope)</p> <ol style="list-style-type: none"> <li>1. Outer diameter 4 mm diameter</li> <li>2. Length should be between 180-185 mm</li> <li>3. Vision 30 degree with 105 degree wide angle</li> <li>4. Fibreoptic light transmission incorporated</li> <li>5. It should offer full O ring illumination</li> <li>6. High depth of focu high speed round lens with high magnification with sapphire glass cover</li> <li>7. Stianless steel body</li> <li>8. Tube laser welded to body welded</li> <li>9. The scope should be autoclavable</li> <li>10. Low vignetty</li> <li>11. The scope should be autoclavable</li> </ol>
2	Vessel-Seal-Cautery	01	<ol style="list-style-type: none"> <li>1. It should have Microprocessor controlled tissue technology.</li> <li>2. It should have pure sinusoidal output waveform in Hf power output in cut and bipolar modes</li> <li>3. It should adjust power level automatically depending on tissue type.</li> <li>4. It Should have Combo Generator Technology having conventional ESU and Vessel Sealing Sysem with Saline Plasma Bipolar Module upgradable.</li> <li>5. It should be useful for conventional underwater procedures.</li> <li>6. It should give an alarm in case of Sealing cable is broken.</li> <li>7. It should have Dedicated Vessel Sealer output with Five different Sealing level programs.</li> <li>8. It should have the capacity to seal the tissue vessel up to</li> </ol>

